

WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	2ZLF
SITE REFERENCE:	
SITE ADDRESS:	Downing Rd. Derby
POST CODE:	DE21 6HA
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	16 x 3

REPORT NO.:	4538886
INDICATOR TYPE:	I30
INDICATOR SERIAL NO.:	04F866408
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	Overcast

REPEATABILITY TEST

Direction of Travel →
←

Indicator Reading: 43,840 kg
Weighbridge Viewed from Indicator Position
Indicator Reading: 43,860 kg

Error Allowed
+/-3dvs or +/-60kg
Pass

END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg) 19,620

Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	End	Middle	End	Zero
3,840		19,620	19,620	19,600	0
	ERROR (e)	error +/-2dvs or 40kg		error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-1dvs or 20kg
24,400	24,240	24,240	0.0	Pass		error +/-2dvs or 40kg
Substitute Load	23,460	23,460	0.0	Pass		error +/-2dvs or 40kg
28,600	28,460	28,460	0.0	Pass	5,000	error +/-2dvs or 40kg
38,600	38,460	38,460	0.0	Pass	10,000	error +/-2dvs or 40kg
43,600	43,460	43,440	-1.0	Pass	5,000	error +/-3dvs or 60kg
44,000	43,860	43,840	-1.0	Pass	400	error +/-3dvs or 60kg

MAINT LOG UPDATED: No

MAXIMUM RANGE TESTED: No
TEST EQUIPMENT USED: Test Weights Traceable to National Standard
CALIBRATION DATE: 16 August 2019

AUTHORISED ENGINEER: Kevin Delve
CERTIFICATE NUMBER: TR10605/TR10606
NEXT CALIBRATION DATE: 21 September 2019

Comments

We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction

Print name	N TIPPIN	Date	16/08/2019
Signature	<i>N TIPPIN</i>	Engineer's Signature	<i>[Signature]</i>

We Certify that the Weighbridge Calibration check complies to the current Weights & Measures criteria and the calibration procedure is carried out in accordance with the Weighing Federation's Code of practice.

(M3LH)