

**VERIFICATION CERTIFICATE – No: 6425**

<b>CUSTOMER:</b> 2ZLF LIMITED	<b>TYPE APPROVAL CERTIFICATE No:</b> LNE-23948 Rev.5
<b>SITE ADDRESS:</b> West Meadows Industrial Estate	<b>INDICATOR TYPE:</b> X112-A (X112-B i30)
Downing Road	<b>INDICATOR TEST CERTIFICATE No:</b> LNE 23273 Rev.5
Derby	<b>INDICATOR SERIAL No:</b> 04F866408
DE21 6HA	<b>MODULE TEST CERTIFICATE No:</b> N/A
<b>MANUFACTURER:</b> Shering	<b>MAXIMUM CAPACITY:</b> 50,000 kg
<b>WEIGHBRIDGE TYPE:</b> Steel Surface Mount	<b>DIVISION (e):</b> 20 kg
<b>PLATFORM SIZE:</b> 15m x 3m	<b>LOADCELL TEST CERTIFICATE No:</b> LNE 17369 Rev 2
<b>EST. PLATFORM WEIGHT:</b> 10,270 kg	<b>PRINTER SERIAL No:</b> N/A
<b>WEIGHBRIDGE LOCATION:</b> Yard Weighbridge	<b>TARE FACILITY:</b> Disabled
<b>CUSTOMER EMAIL ADDRESS:</b> steve.hodges@2zlf.co.uk	
<b>INITIAL VERIFICATION:</b> <input type="checkbox"/>	<b>RE-VERIFICATION:</b> <input checked="" type="checkbox"/>

**Repeatability Test - MPE 0.3e**

	Indication (HR)	Indication (HR)	Indication (HR)
Loaded WTU	44050	44052	44048
	PASS	✓	FAILED

**Accuracy of Zero, Linearity/Hysteresis, & Comparison Tests**

Approximate Test Interval	MPE	Actual Load	Indication Up (HR)	Error	Corrective Error	Indication Down (HR)	Error	Corrective Error	Comparison				
Zero	0.25	0	0	0e		0	0e						
2e	0.5	40	40	0e		40	0e						
20e	0.5	440	440	0e		440	0e						
500e	1.0	10040	10040	0e		10038	-0.1e						
1000e	1.0	20040	20042	0.1e									
Substitute Load	1.0	19328	19330	0.1e		19328	0e		Yes				
1250e	1.0	24328	24332	0.2e		24330	0.1e						
1500e	1.0	29328	29336	0.4e		29334	0.3e						
2000e	1.0	39328	39340	0.6e		39338	0.5e						
Substitute Load	1.0	38632	38644	0.6e		38638	0.3e						
2200e	1.5	44032	44048	0.8e									
<table border="0"> <tr> <td style="border: 1px solid black;">PASSED</td> <td style="border: 1px solid black;">✓</td> <td style="border: 1px solid black;"> </td> <td style="border: 1px solid black;">FAILED</td> </tr> </table>										PASSED	✓		FAILED
PASSED	✓		FAILED										

**Loadcell Positions**

2	3	4
1	6	5
DWI		

FLT Substitute Load  
**4,304 kg**

**Comparison Test**


Printer	<input checked="" type="checkbox"/>
Remote Display	<input checked="" type="checkbox"/>
PC	<input checked="" type="checkbox"/>
Other	
Pass	<input checked="" type="checkbox"/> Fail

**Eccentric Load Test – Up to 500e MPE = 0.5e, over 500e MPE = 1.0e**

Position	1	2	3	4	5	6	7	8	9	10				
Test Load	10040	10040	10040	10040	10040	10040								
Indication (HR)	10044	10040	10040	10038	10036	10040								
Error	0.2e	0e	0e	-0.1e	-0.2e	0e								
Corrective Error														
<table border="0"> <tr> <td style="border: 1px solid black;">PASSED</td> <td style="border: 1px solid black;">✓</td> <td style="border: 1px solid black;"> </td> <td style="border: 1px solid black;">FAILED</td> </tr> </table>											PASSED	✓		FAILED
PASSED	✓		FAILED											

**ENVIRONMENTAL CONDITIONS:** Light Rain and Windy  
**TEST WEIGHTS USED:** Precia-Molen UK Ltd  
**VERIFICATION DATE:** 01/10/2021

**AUTHORISED ENGINEER:** Mark Spiers  
**CERTIFICATE NO:** TR 11135 & TR 11136  
**NEXT CHECK DATE:** 01/10/2022

Signature: .....  ..... Date: ...13/10/2021.....

*This Verification Certificate complies with the requirements of NAWI directive 2014/31/EU and EN45501:2015. The tests are also carried out in compliance with the UK Weighing Federation Code of Practice for Non Automatic Weighing Instruments.  
 Verification Certificate CP 81015.01. – Issue: V1.1 - 03/01/2021*