

## WEIGHBRIDGE CALIBRATION REPORT - "AS FOUND"

CUSTOMER:	2ZLF
SITE ADDRESS:	Downing Rd.
	Derby
POST CODE:	DE21 6HA
WEIGHBRIDGE LOCATION:	Main Weighbridge
WEIGHBRIDGE SIZE:	16 x 3

SAP REPORT NO.:	4812141
INDICATOR TYPE:	I30
INDICATOR SERIAL NO.:	04F866408
MAXIMUM CAPACITY (kg):	50,000
DIVISION 'e' (kg):	20
Weather Conditions	Sunny

### REPEATABILITY TEST

Direction of Travel	→	Indicator Reading: <b>44,180 kg</b>	Error Allowed +/-2dvs or +/-40kg <b>Pass</b>
	←	Weighbridge Viewed from Indicator Position Indicator Reading: <b>44,180 kg</b>	

### END - MIDDLE - END - ZERO TEST

APPLIED LOAD (kg)	15,000
-------------------	--------

		End	Middle	End	Zero
Fork Lift Truck Weight (kg)	INDICATOR READING (kg)	15,000	15,000	15,000	0
<b>4,700</b>	ERROR (e)	error +/-2dvs or 40kg	error +/-2dvs or 40kg	error +/-2dvs or 40kg	error +/-1dvs or 20kg
	PASS/FAIL	Pass	Pass	Pass	Pass

### RANGE TEST

Target APPLIED LOAD (kg)	Actual APPLIED LOAD (kg)	DISPLAYED READING	ERROR (e)	PASS/FAIL	Added weight (kg)	Allowed Error
0	0	0	0.0	Pass		error +/-1dvs or 20kg
5,000	5,000	5,000	0.0	Pass		error +/-1dvs or 20kg
10,000	10,000	10,000	0.0	Pass		error +/-1dvs or 20kg
20,000	20,000	20,000	0.0	Pass		error +/-2dvs or 40kg
Substitute Load	19,480	19,480	0.0	Pass		error +/-2dvs or 40kg
25,000	24,480	24,480	0.0	Pass	5,000	error +/-2dvs or 40kg
35,000	34,480	34,480	0.0	Pass	10,000	error +/-2dvs or 40kg
40,000	39,480	39,480	0.0	Pass	5,000	error +/-2dvs or 40kg
Substitute Load	39,180	39,180	0.0	Pass	0	error +/-2dvs or 40kg
44,000	44,180	44,180	0.0	Pass	5000	error +/-3dvs or 60kg

MAINT LOG UPDATED:

MAXIMUM RANGE TESTED:

TEST EQUIPMENT USED: Test Weights Traceable to National Standards

CALIBRATION DATE: 13 September 2022

AUTHORISED ENGINEER: Kevin Delve

TEST WEIGHT CERTIFICATE NUMBER: TR11411/TR11412

NEXT CALIBRATION DATE: 12 September 2023

Comments			
<i>We hereby approve/confirm the above is a true record of the work undertaken which has been carried out to our satisfaction</i>			

Name		Date	13/09/2022
Customer Signature		Engineer's Signature	

*We Certify that the calibration check complies to current Weights & Measures criteria and the calibration procedure is carried out in accordance with the UK Weighing Federation's Code of practice.*